



Optical Coherence Tomography
RS-3000 Advance




THE ART OF EYE CARE

US EDITION


Wide Area and High Definition OCT with SLO Tracing


12 mm wide horizontal scan available with the RS-3000 Advance allows detailed observation of the vitreous body, retina, and choroid from the macula to optic disc in a single image.



Selectable OCT Sensitivity


Selecting the OCT sensitivity based on ocular pathology allows image capture with higher definition or at high speed. Ultra fine, fine, and regular sensitivities are available for the RS-3000 Advance. Ultra fine and fine sensitivities are used to capture high definition images and regular sensitivity is used to capture images at high speed.






Wide Area Scan 9 x 9


A 9 x 9 mm* wide area image centering around the macula can be captured with the RS-3000 Advance.




ILM-RPE / BM



ILM-IPL / INL



ILM-NFL / GCL




NFL/GCL-IPL / INL

Tracing HD plus


The tracing HD plus function in the RS-3000 Advance traces involuntary eye movements to maintain the same scan location on the SLO image for accurate image capture. This function allows accurate averaging of up to 120 images.

Macula multi (cross)

The macula multi scan pattern captures 5 cross-sectional images each in the X and Y directions. High-quality images are easily obtained with the tracing HD plus function.



Capture screen



OCT image with averaging of 10* images

*The maximum number of images that could be averaged with previous software.


OCT image with averaging of 30 images

Macula radial and disc radial

The macula radial and disc radial scan patterns capture 6 or 12 radial cross-sectional images centered on the macula and optic disc respectively. The tracing HD plus function ensures the scan is centered on the targeted region.




OCT image with averaging of 30 images




OCT image with averaging of 120 images

HD checker

The HD checker function in the RS-3000 Advance displays the image during averaging and allows an operator to check and finish capturing prior to reaching the number for averaging set by an operator if sufficient image quality is obtained.



OCT image with averaging of 30 images




OCT image with averaging of 100 images

Torsion Eye Tracer (TET)

The TET incorporated in the RS-3000 Advance ensures accurate image capture by utilizing fundus information from the high definition SLO image. The three functions, positioning, tracing, and auto shot allow accurate image capture of the targeted region. Ocular cyclotorsion is traced via the torsion correction feature added to the tracing function.

Torsion correction

The torsion correction function ensures the scan is oriented at the right angle even in cases of ocular cyclotorsion and fundus tilt due to head movement or incorrect positioning on the chinrest and forehead rest.



Positioning



The positioning function briefly provides a still live SLO image in order to easily locate the target of interest on the fundus.

Tracing

The tracing function automatically traces the fundus after positioning is completed. It ensures the scan is centered on the target.

Auto shot

The auto shot function enables automated image capture when the scan is centered on the target and oriented at the right angle. It avoids capturing images in mid-blink or images with incorrect fixation.




Follow-up Image Capture

The follow-up image capture function in the RS-3000 Advance performs positioning based on the previously captured baseline data, and then tracing and auto shot. It provides ease-of-use and high reproducibility of the image capture for follow-up examination.

The baseline data can be easily registered with one button.



Each baseline image can be displayed in the thumbnail view.



The previously captured position is automatically traced by selecting the registered baseline.




Retina Analysis

Retinal and choroidal modes are available for the RS-3000 Advance. The choroidal mode allows a more detailed examination of the choroid.


Macula line with 12 mm horizontal scan

The macula line scan pattern captures a cross-sectional image at a user designated position. The 12 mm horizontal scan of the RS-3000 Advance allows observation of a wide area from the macula to the optic disc in a single image.




Choroidal OCT image (EDI-OCT)

Choroidal mode in the RS-3000 Advance allows capture of highly reflective choroidal images by reversing the image.




Enhanced image

The enhanced image function allows to adjust bright intensity of an image to enhance details.




Flexible cross scan



The flexible cross scan mode* in the RS-3000 Advance allows free placement of the scan position within a capturing window by shifting the crossing point of the scan pattern lines. This function is useful for capturing an image of pathology that is distant from the center of the SLO image.


*The flexible cross scan mode is available for the macula cross and macula multi (cross) scan patterns.




Select and Rescan mode (SR mode)

The select and rescan mode in the RS-3000 Advance allows capture of an entire image of the retina with the macula map scan pattern and select a cross-sectional OCT image with the location of lesion from up to 256 images based on user preference. Cross-sectional OCT images can be reacquired on the selected region with the tracing HD plus function. The select and rescan mode is useful in efficiently obtaining a high-quality image of a region of interest.

Macula map scan



Select an image and then confirm with the select and rescan mode button.



Macula line scan starts on the selected region.

Macula examination

Macula multi (cross)




The macula multi scan pattern enables acquisition of 5 cross-sectional images each in X and Y directions. The appropriate image for diagnosis can be selected from the 10 images.



Macula map



The macula map scan pattern captures up to a 9 x 9 mm area and provides a color-coded map, thickness analysis, surface, and retinal layers.



Analysis chart

Glaucoma Analysis

Wide area scan 9 x 9

Wide area images of 9 x 9 mm centered on the macula can be acquired with the RS-3000 Advance.



Macula map



The glaucoma analysis provides the [NFL+GCL+IPL] analysis, which supplements clinical work-up for the early detection of optic nerve fiber layer defects. The 9 x 9 mm wide area map enables analysis of the [NFL+GCL+IPL] in the peripheral retina.

Analysis display switching tab

[NFL+GCL+IPL]
Color-coded thickness map (12 x 9 mm) of [NFL+GCL+IPL] layers (ILM to IPL / INL) overlaid on SLO


Cross-sectional image display

Analysis charts
(Superior / Inferior pole, GChart)


Analysis charts of average thickness of each sector surrounding the macula

[NFL+GCL+IPL]


The [NFL+GCL+IPL] are layers composed of Nerve Fiber Layer (NFL), Ganglion Cell Layer (GCL), and Inner Plexiform Layer (IPL).



NFL
GCL
IPL



Color fundus photography taken with another device



[NFL+GCL+IPL] thickness map

Disc map




The disc map scan pattern captures an image centered on the disc and provides data for comprehensive disc analysis.

RNFL thickness map
Color-coded thickness map of RNFL layer (ILM to NFL / GCL)

SLO image*
SLO image showing optic disc

TSNIT graph
Graph showing thickness from ILM to NFL / GCL on disc circle

OCT image of disc circle



Analysis table
Table of optic disc analysis
C / D ratio (horizontal)
C / D ratio (vertical)
R / D ratio (minimum)
R / D ratio (angle)
Disc area (mm²)
Cup area (mm²)

Overall tab displaying both right and left eyes


SLO image*
SLO image showing scanned area with color-coded thickness map of user selected layers

Analysis charts
Analysis charts indicating average thickness of Whole, S / I (2-sector), TSNIT (4-sector), and Clock Hour (12-sector)

Disc circle




The disc circle scan pattern in the RS-3000 Advance captures an image of circle in 3.45 mm diameter centering on the disc and allows RNFL.



Disc radial



The disc radial scan pattern in the RS-3000 Advance captures 6 or 12 radial cross-sectional images centered on the disc and allows observation of disc shape symmetry.



Multifunctional Follow-up

The multifunctional follow-up allows analysis of all the data obtained with the OCT and detailed observation of chronological change in retinal thickness and status. This function displays progression of pathology over the short term, intermediate- and long-term together with a numerical value obtained from RS-3000 Advance and other examinations such as intraocular pressure and visual field, which provide clinical information for guiding treatment.

Progression mode

The progression mode performs data analysis based on the data captured up to 50 times and displays chronological change in retinal thickness with various maps, charts, and graphs for trend analysis.

Analysis display tab
Progression: Macula / Glaucoma / Custom
Comparison: Macula / Glaucoma

Parameter of the captured data
Date: Capture date (age)
SQL: SLO quality / SSI: OCT quality
Follow-up: Time from the date of baseline data

Map / Data display switching buttons
Anterior segment
SLO fundus image*
Scan position correction
Color map

Selection of various parameter sectors
Up to 6 different analyses can be specified.

Graph indicating trend of retinal thickness

Event registration
Treatment record such as start of medication or surgery can be registered and displayed.

Custom registration
Record of the numerical value such as intraocular pressure and visual acuity can be registered and displayed.

Graph analysis of retinal thickness trend
Slope and p-value of each graph

Event	Slope(µm)	pValue
1	+3.33 µm	...
2	-3.45 µm	< 0.01
3	+1.10 µm	...
4	-4.38 µm	< 0.01
5	-8.34 µm	< 0.01
6	-3.38 µm	< 0.01

Comparison mode

The comparison mode displays two images selected by the user for comparison and analysis of retinal thickness.

Map / Data display switching buttons
Anterior segment
SLO fundus image*
Scan position correction

Graph indicating change in thickness


Comparison data

Customized Report

The layout of the reports can be customized and the data from separate reports of each scan pattern can be summarized in a single report to avoid printing multiple pages. The report setting can be titled such as glaucoma, macular disease, and screening based on user preference.

Glaucoma

Only the necessary images and analysis results obtained with the macula map (both eyes), disc map (both eyes) are summarized in a report.



Macula map

- ILM to IPL / INL color map
- GChart, S / I analysis chart

Disc map

- ILM to NFL / GCL color map
- Various analysis charts
- TSNIT graph

Combo release mode

The combo release mode combines scan patterns and facilitates an examination requiring several scan patterns. The scan patterns and their order can be user specified. The scan pattern selected for combo release mode can be preset and reflected on the report.



Default setting

Macula disease



Macula multi



Macula map X-Y

Glaucoma



Macula map Y-X



Disc map Y-X

Macula (one eye)

Macula multi (cross)

Macula map

Sample of customized report

Macula (both eyes)

Macula cross*


Macula map

Sample of customized report

Anterior Segment Analysis

The optional anterior segment module enables observation and analyses of the anterior segment.


Angle measurement




- ACA
Angle between posterior corneal surface and iris surface
- AOD500 (AOD750)
Distance between iris and a point 500 μm (or 750 μm) away from scleral spur on posterior corneal surface
- TISA500 (TISA750)
Area circumscribed with AOD500 (or AOD750) line, posterior corneal surface, line drawn from scleral spur in parallel with AOD line, and iris surface



Cornea measurement



- Corneal thickness
Corneal thickness of apex and user's preferred sites
- Corneal thickness map
Map indicating corneal thickness measured in radial directions



Anterior segment adaptor



NAVIS-EX is an image filing software, which networks the RS-3000 Advance and other NIDEK fundus imaging devices, the AFC-330 and F-10.

Non-mydratric auto fundus camera

AFC-330



Examination room

LAN connection


Diagnostic room



NAVIS-EX allows viewing and data analysis of captured images in the diagnostic room.




Another server may be necessary depending on the network setup.

Model	RS-3000 Advance
Fundus surface imaging	 <p>SLO (12 fps frame rate) 40° x 30° angle of view</p>
Scan speed	Max. 53,000 A-scans / s
OCT sensitivity	Regular, Fine, Ultra fine
Scan pattern (retina)	Macula line (scan angle changeable by 1°) Macula cross Macula map (with cross scan / without cross scan) Macula multi (X-Y: 5 x 5) Macula radial (6 lines / 12 lines) Disc circle Disc map Disc radial (6 lines / 12 lines)
Scan pattern (cornea) with optional anterior segment module	Cornea line Cornea cross Cornea radial (6 lines / 12 lines) ACA line
Image averaging	Max. 120 images
Choroidal mode	Available
Torsion eye tracer	Available
Follow-up tracing	Available
Follow-up analysis	Available
Tracing HD plus	Available
HD checker	Available
Flexible cross scan	Available
Select and rescan mode	Available
Auto shot (for follow-up image capture)	Available
Internal fixation target	Cross shape (laser)
PC monitor	21"

RS-3000 Advance Specifications

Model	RS-3000 Advance
OCT scanning	
Principle	Spectral domain OCT
OCT resolution	Optical Z: 7 μ m, X-Y: 20 μ m Digital Z: 4 μ m, X-Y: 3 μ m
Scan range	X: 3 to 12 mm (12 mm for line scan only) Y: 3 to 9 mm Z: 2.1 mm
OCT light source	SLD, 880 nm
Scan speed	Max. 53,000 A-scans / s
Acquisition time of 3-D image	1.6 s in regular mode
Internal fixation lamp	637 nm
External fixation lamp	630 / 565 nm
Auto alignment	Z direction
Minimum pupil diameter	ϕ 2.5 mm
Focus adjustment range	-15 to +10 D (VD=12 mm)
Working distance	35.5 mm
Software analysis	Segmentation of 6+1 retinal layers Macular thickness map RNFL thickness map [NFL+GCL+IPL] analysis Optic nerve analysis Follow-up analysis
Fundus surface imaging	
Principle	Confocal scanning laser ophthalmoscope (SLO light source: 785 nm)
Angle of view	40° x 30° (zoom: 20° x 15°)
PC networking	Available
Display	Tiltable 8.4-inch color LCD
Power supply	AC 100, 120, 230 V 50 / 60 Hz
Power consumption	300 VA
Maximum power output (transformer)	1,000 VA
Dimensions / Mass	380 (W) x 524 (D) x 499 to 531 (H) mm / 34 kg 15.0 (W) x 20.6 (D) x 19.6 to 20.9 (H)" / 75 lbs.

Specifications may vary depending on circumstances in each country. Specifications and design are subject to change without notice.



Anterior segment module (optional)	
Software analysis	Corneal thickness measurement Corneal thickness map Angle measurement

Motorized optical table (optional)	
Dimensions / Mass	639 (W) x 472 (D) x 600 to 850 (H) mm / 28 kg 25.2 (W) x 18.6 (D) x 23.6 to 33.5 (H)" / 62 lbs.
Power supply	AC 100 V (available from the transformer) 50 / 60 Hz
Power consumption	150 W

PC rack (optional)	
Dimensions / Mass	620 (W) x 460 (D) x 700 (H) mm / 29 kg 24.4 (W) x 18.1 (D) x 27.6 (H)" / 64 lbs.

Listed features in this brochure are intended for US practitioners. Specifications may vary depending on circumstances in each country. Specifications and design are subject to change without notice.



RS-3000 Advance



HEAD OFFICE
34-14 Maehama, Hiroishi
Gamagori, Aichi 443-0038, Japan
Telephone: +81-533-67-6611
Facsimile: +81-533-67-6610
URL: <http://www.nidek.co.jp>

[Manufacturer]

TOKYO OFFICE (International Div.)
3F Sumitomo Fudosan Hongo Bldg.,
3-22-5 Hongo, Bunkyo-ku, Tokyo
113-0033, Japan
Telephone: +81-3-5844-2641
Facsimile: +81-3-5844-2642
URL: <http://www.nidek.com>

NIDEK INC.
47651 Westinghouse Drive
Fremont, CA 94539, U.S.A.
Telephone: +1-510-226-5700
: +1-800-223-9044 (US only)
Facsimile: +1-510-226-5750
URL: <http://usa.nidek.com>

NIDEK S.A.
Europarc
13, rue Auguste Perret
94042 Créteil, France
Telephone: +33-1-49 80 97 97
Facsimile: +33-1-49 80 32 08
URL: <http://www.nidek.fr>

NIDEK TECHNOLOGIES Srl
Via dell'Artigianato, 6 / A
35020 Albignasego (Padova), Italy
Telephone: +39 049 8629200 / 8626399
Facsimile: +39 049 8626824
URL: <http://www.nidektechnologies.it>

